User Manual



FAST Chair (Removable Stair Track)

CE

Ferno Slovakia s.r.o.
913 07 Bosaca 893 Slovakia
Tel: + 421 (0) 32 7708 010, sales.sk@ferno.com

July 2024, ver. 08


Pub. No. 50-8110-011_08

Table of Contents

Disclaimer 4
Technical specification: 4
Using Patient Restraints 6
4-Point Restraints (accessory) 7
1. Quick set-up guide 8
1.1 Unfold chair 8
1.2. Adjust handle 9
1.3. Unfold & Fold track 10
1.4. Telescopic carry handles 11
1.5 Attaching the Track (Chair Occupied) 12
1.6 Attaching the Track (Chair NOT Occupied) 13
1.7 Deploying the Track for Use 13
1.8 Stowing the Track 14
1.9 Removing the Track 14
1.10. Fold chair 15
1.11. Foot rest 16
2. Intended use 17
3. Safety Precautions 17
3.1. Operator Skills 17
3.2. Operator Training 17
3.3. Warning 18
4. FAST Chair Operation 19
4.1. Folding and Unfolding the Chair 19
4.2. Height Positional Upper Handle 19
4.3. Extending Carrying Handles 19
4.4. Patient Restraint 20
4.5. FAST Chair Wheels 20
4.6. FAST Chair 20
4.7. Checking the FAST Chair 21
5. Operating Procedure 22
5.1. General Information 22

5.2. Moving the Chair 22
5.3. Using the Chair on Level Ground 23
5.4. Gliding the Chair Downstairs 23
6. Care of the FAST Chair 25
6.1. Disinfection 25
6.2. Cleaning 25
7. Using additional help 26
8. Maintenance 27
9. Warranty and Service 27

Disclaimer



This manual contains general instructions for the use, operation and care of this chair. The instructions are not all-inclusive. Safe and proper use of this chair is solely at the discretion of the user. Safety information is included as a service to the user. All other safety measures taken by the user should be within and under consideration of applicable regulations. It is recommended that training on the proper use of this chair be provided before using this chair in an actual situation.

Retain this manual for future reference. Include it with the chair in the event of transfer to new users. Additional free copies are available upon request from FERNO Manufacturing.



Technical specification:

Depth of seat	***450 mm***
Height of seat from floor	***475 mm***
Width of seat	***415 mm***
Front wheels /dia/	***125 mm***
Rear wheels /dia/	***150 mm***
Height of top handle max	***580 mm***
Overall length	***695 mm***
Overall height	***1000 mm***
Overall width	***516 mm***
Weight /without track/	***11,2 kg***
Weight /with track/	***15 kg***
Load capacity	***250 kg***

TIP: Regular cleaning helps reduce the risk of transmitting disease and enables the equipment to function at its optimum

⚠ WARNING

Untrained operators can cause injury or be injured. Permit only trained personnel to operate the chair.

Improper use of the chair can cause injury. Use the chair only for the purpose described in this manual.

Improperly attaching restraints can allow seat and back panels to dislodge, resulting in injury. Restraints must capture panel and chair frame.

Improper charging can cause injury or damage. Use only the Ferno-approved battery with the chair, and charge only with the Ferno-approved battery charger.

Using the chair with the track system unlocked can cause injury. Verify that the track system is locked before transporting the patient.

Lubricating the track system can result in injury to the patient and/or operators. Never lubricate the tracks or belts.

Improper operation can cause injury. Operate the chair only as described in this manual.

An unattended patient can be injured. Stay with the patient at all times.

An unrestrained patient can fall off the chair and be injured. Use restraints to secure the patient on the chair.

Helpers can cause injury or be injured. Maintain control of the chair, operate the controls, and direct all helpers.

Improper maintenance can cause injury. Maintain the chair only as described in this manual.

Attaching improper items to the chair can cause injury. Use only Ferno-approved items on the chair.

Improper parts and service can cause injury. Use only Ferno parts and Ferno-approved service on the chair.

Modifying the chair can cause injury and damage. Use the chair only as designed by Ferno.



SPARE PARTS:

Item	Part number	Description	QNT.
1	65-4119-001	ASSY LOCKING TELESCOPE	1
2	65-0402-001	REMOVABLE MANUAL STAIR TRACK	1
3	65-4113-001	ASSY HANDLER INT OVAL PROFIL	2
4	65-4155-001	ASSY LOCKING FOOTS	1
5	75-0004-044	FRONT CHAIR WHEEL 125 MM	2
6	75-0004-043	REAR CHAIR WHEEL 150 MM	2
7	65-4118-001	SEAT VINYL FS CHAIR	1
8	65-4126-001	HEAD SUPPORT	1
9	65-0209-002	FOOT RESTRAINT, 50 MM (S-240\242)	1
10	65-0212-005	RESTRAINT, 1 PC 180 CM, BLAK PP, PLASTIC BUCKLE	2
11	50-4158-001	PIPE BACK WHEEL	1

ACCESSORY:

Item	Part number	Description	QNT.
12	65-4122-001	HANDLE, ASSEMBLY	1
13	65-4130-001	ASSY COMFORT FOOTBAR	1
14	65-4123-001	MASH BASKET FOR OXYGEN BOTTLE	1
15	62-4400-001	4-POINT RESTRAINT	1
16	65-4128-001	ARMREST KIT	1
17	65-4137-001	FRONT WHEELS LOCKING KIT	1

The CE mark on this product demonstrates conformity with the requirements of the MDR 2017/745 for Medical Devices and EN 1865.

CE

Using Patient Restraints

Use the lap and foot restraints to secure Patient into the Chair. Assess Patient condition to determine optimum strapping procedure.

• Chest/Hips Restraint: The chest restraint can also be used to secure the chair closed for stowage. Fasten the chest\hips restraint using plastic male/female buckling system. Pull the loose free end of the webbing strap to tighten the restraint.





• Foot Restrain: Once the patient's feet are comfortably placed on the integral footrest, the loose foot restraint can be fastened to secure the feet. Fasten foot strap using plastic male/female buckling system. Pull the loose free end of the webbing strap to tighten the restraint. Remember that the foot restraint also protects the operator when moving the Chair on stairs.



• Releasing Restraints: To release foot, chest and hips restraints, squeeze the two sides of the tang on the buckle inwards at the same time to pull the buckle apart.





4-Point Restraints (accessory)

1. To attach the seat restraint, lay the restraint on the top of the seat panel and feed the male and female ends of the restraint down through the slots.
2. Wrap the restraint around the aluminum chair frame on each side of the chair, capturing both the panel and the frame.
3. Buckle the restraint and adjust the length.
4. Repeat Steps 1-3 to attach the back-panel restraint.





1. Quick set-up guide

1.1 Unfold chair

For the purpose of this guide, the FAST Chair will be referred to as the Chair and the person being transported on the Chair will be referred to as the patient.

The Chair will normally be left in the folded condition and the first task will be to unfold and make the Chair safe for transporting a person.

Unbuckle the chest strap securing the chair and fold the chair out into the seated position.

Completely unfold the chair by holding the handle upright and firmly pushing down on the seat until an audible click is heard, this is the locking pins sliding into place, visually confirm this before seating a patient.





⚠ WARNING
Do not pull down the footrest until patient is seated!!

1.2. Adjust handle

The chair is fitted with an adjustable height handle. To adjust the handle, the operator must pull up on the red cable under the handle. This disengages the handle to slide up or down. When the desired position is reached the operator must release the cable.





Pull the red cable under the handle upward and slide the handle up or down to its desired position.



Final open position

Once the patient has been seated comfortably in the chair pull down the foot rest and place their feet on to it. Fasten all restraining straps to secure the patient.





1.3. Unfold & Fold track

The track aids descent on stairs, do not unfold the track if the chair is not needed to descend stairs. To fold the track into the back of the chair push the lever downwards as shown, until track is flush with the back of the chair and locks in to position.





To unfold the track from the rear of chair pull the bar fitted to the track until the track locks in to position.





1.4. Telescopic carry handles

The chair is fitted with telescopic carrying handles to aid carrying the chair down stairs. To change position of the handles press button and pull/push handles to extend or shorten. To aid carrying the chair down/upstairs spill Rear locking lift handle. After using come back both handles to the starting position.






1.5 Attaching the Track (Chair Occupied)

The chair features an attachable track to aid operators when descending stairs.

1. With the Chair in the unfolded position, raise the red release handle at the bottom of the track and locate the lower track lock onto the axle. Ensure the track is locked in place.



Track attached on axle



Locating lower Track lock onto axle

2. Bring the Track upwards towards the chair and push into the chair until an audible click is heard, indicating the striker pins have locked into the push-to-close latches.

3. Always carry out a visual inspection to ensure both the location pins have located and attached the Track securely to the rear of the Chair.



Rotating Track upwards to attach to Chair



Track attached to Chair back via push-to-close latches

1.6 Attaching the Track (Chair NOT Occupied)

When the Chair is not occupied, an alternative method for attaching the Track is recommended.
1. Lay the Track on the floor and lift the Chair into place, locating the axle bar in to the lower Track lock.



Preparing to attach Track to unoccupied Chair

2. Push the Chair axle securely into the lower lock until it is locked in place.
3. Rotate the Track upwards and lock in place as described previously in steps on previous page.



1.7 Deploying the Track for Use

Once the Track are attached to the chair, they may be unfolded for use.
Whilst holding the Chair secure from the rear, pull the red 'Pull to Open' bar forwards and down, until the Track locks into position.



Pulling Track out for use




Chair with Track deployed ready for use

1.8 Stowing the Track

When not in use always ensure the Track is folded away securely to limit the chance of damage or injury.

1. To fold the track into the back of the chair, push the bar with RED 'Push to Close' label downwards as shown, until the track is flush with the back of the chair and locks into position.



Using 'Push to Close' bar to stow Track




Track stowed

1.9 Removing the Track

Prior to folding the chair for storage or using chair as a carry chair, the Track must be removed. When the Chair is occupied with a patient, remove the track using the following procedure:

1. Pull the RED Track release bar down towards the floor to release the locating pins from the two push-to-close latches and detach Track from the top of the Chair.



Releasing Track from push-to-close latches



Releasing the Track from the Chair

1.10. Fold chair



Pull the red cable under the seat outward



Fold the seat up



Continue folding the seat into its folded position



Folded position

1.11. Foot rest

Before folding the chair for storage ensure castors are locked facing forward, the foot rest is up and all handles are not extended.





The foot support is part of this chair. For unfolding is necessary to pull the chair frame down and for folding is necessary to pull the chair frame up.



2. Intended use

For the purpose of this manual, the FAST Chair will be referred to as the Chair and the person being transported on the Chair will be referred to as the patient. The Chair is a handling device designed to transport a patient, in a seated position, on stairs and level ground. The Chair will glide smoothly down a wide range of stair types and can quickly return to a four wheel chair at the end of a run of stairs. When on two or four wheels it can rotate within its own length making it ideal for half landings and similar restricted areas. Where conditions require the chair may also be carried using the anti-slip handle grip areas.

3. Safety Precautions

To ensure safety and satisfaction with your new Chair, all users should read these instructions and the individual portions of recognized emergency care manuals that relate to the application and use of Carrying Chairs. It is recommended that a minimum of two trained operators control a loaded Chair on stairs, they should assess whether additional assistance is required.

3.1. Operator Skills

Minimum recommended skills are as follows:

• The strength, balance, co-ordination and common sense to safely operate the Chair.

• The ability to select the proper equipment and procedures for the circumstances.

• Operators should have been trained in the operation of the Chair.

• The ability to understand and follow the procedures described in this manual.

• Operators should have a working knowledge of emergency patient-handling procedures.

3.2. Operator Training

Minimum recommended training is as follows:

• A training plan approved by an EMS training officer (or equivalent).

• Read and understand the instructions in this manual.

• Practice using the fully-equipped Chair with weight and under simulated conditions.

• Trainees should be tested to verify their understanding of Chair operation.

• Keep training records. Annual refresher training for all Chair operators is recommended.

3.3. Warning

Warning notices indicate a potentially hazardous situation which if not avoided, could result in injury or death.

WARNING

Bleach, phenols and iodine can cause damage. Do not apply products containing these chemicals to the Chair.

Improper lubricants can cause damage. Use light oil to lubricate the Chair.

WARNING

Blood borne Disease Notice -To reduce the risk of exposure to blood borne diseases (such as HIV-1 and hepatitis), when using the Chair, follow the maintenance instructions in this manual.

Rear locking lift handle is intended only for over stairs transport, not for chair turning.

WARNING

NEVER LUBRICATE OR ALLOW LUBRICANTS TO COME INTO CONTACT WITH THE TRACK SYSTEM

WARNING

Unfold the track for use only when it is occupied and only when you need to transfer the patient upstairs or downstairs.

Every three months lubricate the hinge pins with light oil paying particular attention to the cross section bolts.

WARNING

- ✓ **Never exceed the maximum load capacity of 250 Kg.**
- ✓ **Only qualified personnel trained in the use of the Chair should use it.**
- ✓ **For safety purposes, it is essential that patients should always be secured with restraints.**
- ✓ **Never leave a loaded Chair unattended**
- ✓ **An unloaded Chair may tip backwards. If located on an incline, wherever possible face. The Chair down an incline with the castor locks engaged.**
- ✓ **Avoid rapid Chair repositioning – this may cause injury to the patient. Support the Chair while adjusting handles**
- ✓ **Establish regular inspection procedures and responsibilities to ensure proper maintenance as described.**
- ✓ **In cases of cervical, spinal or some fracture injuries, use of the Chair for patient transfer is not recommended.**
- ✓ **Never leave the Chair unsupported with a person on-board as the Chair is designed to tilt back easily.**
- ✓ **Improper maintenance can cause injury and unpredictable operation. Maintain the Chair only as described in this manual.**
- ✓ **Improper parts and service can cause injury and damage. Use only FERNO-approved parts and service.**

4. FAST Chair Operation

- Refer to Quick Set-up Guide

4.1. Folding and Unfolding the Chair

The Chair folds for shipping and storage.

To Unfold

1. Release the chest strap from the Chair and allow the chair to fold down into the seated position.
2. Completely unfold the chair by holing the handle upright and firmly pushing down on the seat until an audible click is heard, this is the locking pins sliding into place, visually confirm this before seating a patient.

To Fold

When folding the chair ensure the following procedure is carried out:

1. Fully lower the Upper Handle (see below).

2. Check that the carrying handles at the front of the chair are not extended.

3. Ensure foot rest is folded up.

4. Ensure the track is folded flush with the back of chair.

5. Ensure the rear wheels are locked and the front castors are facing forward before folding. This allows for easy and safe folding.

6. To release the chair for folding, pull the red cable under the seat outward and fold the chair forward until it is fully collapsed.

7. Stand the chair upright on the back wheels.

8. The wheels can now be unlocked so that the chair may be rolled into position.

9. The Chair chest strap can be used to secure the Chair in the closed position.

4.2. Height Positional Upper Handle

The upper handle can be locked into four different positions to assist operations on stairs and also to allow for height variations of operator and patient. The chair is fitted with an adjustable height handle. For the handle adjustment into other requested position it is necessary to pull and rotate on 90° degrees of the groove pin at the right side handle in the order to lock off at the right side handle. Thereafter it is necessary to lock off left groove pin at the left side by pulling and holding in unmeshed position. You have to adjust the handle by movement down or movement up at pin pulling till reaching of requested position. After groove pin locking off at the left side, it is necessary to lock the pin at the right side handle. Having to be sure about correct pin fitting (locking) resulting the correct locking handle position.

4.3. Extending Carrying Handles

The telescopic carrying handles are located at the front of the chair just above the castor wheels and are to be used in conjunction with the Track Wheels. When using the Track Wheels extend the handles to suit the height of the operator.

The handles are easily extended or shortened by pressing down on the buttons located on the handles with the thumb and pulling or pushing the handles to extend or shorten. Ensure the handles are locked before attempting to lift or carry the chair.

4.4. Patient Restraint

Use the lap and foot restraints to secure patient into the Chair. Assess patient condition to determine optimum strapping procedure.

1. Chest Restraint - The chest restraint is also used to secure the chair closed for stowage. Fasten the chest restraint using plastic male/female buckle system. Pull the loose end free end of the webbing strap to tighten the restraint.

2. Foot Restraint-Once the patient's feet are comfortably placed on the integral footrest, the loose foot restraint can be fastened to secure the feet. Fasten foot strap using plastic male/female buckle system. Pull the loose free end of the webbing strap to tighten.

3. Releasing Restraints -To release foot and lap restraints, squeeze the two sides of the tang on the buckle inward at the same time and pull the buckle apart.

4.5. FAST Chair Wheels

The Chair is fitted with two large wheels with brakes to the rear and two swiveling castors below the foot rest.

• The Chair may be wheeled using all four wheels where the front swivel castors provide excellent maneuverability, alternatively, should conditions require, the chair may be tilted back and run on the two larger wheels. ALWAYS inform the patient of your intentions.

• To apply the brakes on the rear wheels, press down on the brake pedal above the wheels. To release the brake raise the brake pedal with your foot.

4.6. FAST Chair

In addition to the wheels the Chair is fitted with a track system to allow the operators to glide the chair down stairs instead of lifting and carrying.

Guidelines for using the track system:

• Using the Chair on stairs requires a minimum of two trained operators. FERNO Manufacturing recommends using a third person as a "spotter" (see, using the Chair on Stairs).
• Verify that the track is fully opened and locked into place before using it.
• Verify that the telescopic handles are locked before lifting.
• Never lubricate track belts. Lubricated track belts can perform unpredictably, resulting in injury to the patient and/or operators.
• Moisture, water, snow, ice or debris on or between the track and belts can cause irregular track-belt performance that results in sudden changes in the weight operators must support. Make sure the track and track belts are clean and dry before using the Chair on stairs.
• Always clear stairs of debris etc., which could interfere with the operation of the Chair or impede the operators.

Opening the Track:

1. Grasp the top of the chair with one hand and with the other grasp the track release bar (red labelled) located near the top of the tracks and firmly pull it back until the track locks into the fully extended position.
2. Verify that the lock is engaged by trying to push the track closed. If the lock is fully engaged the track will not move.

Closing the Track:

With one hand, grasp the track release bar (red labelled) located near the top of the tracks and with the other hand push down on the track closing handle located on the right side of the tracks. Pushing the track closed with the track release bar will assist this action.

4.7. Checking the FAST Chair

The following checks are required each day before the use of the Chair:

• Overall cleanliness of the Chair.

• Check that there are no bent or damaged parts, missing fasteners, etc.

• Check that the wheels rotate freely with no excess movement or damage.

• Check that the Chair unfolds and locks securely in the seating position for normal use.

• Check that the track system opens and closes correctly, examine the tracks for damage.

• Check the upholstery for cracks, splits or any loose fabric or stitching, etc.

• Check that the upholstery is correctly secured to the frame.

If in any doubt then contact a FERNO Field care Service Engineer.

5. Operating Procedure

5.1. General Information

IMPORTANT:

Always inform the patient before a manoeuvre.

Chair operation requires a minimum of:

• Two trained operators when the Chair is being used to transport a patient on stairs.

• One operator only, can safely use the Chair on flat surfaces including single steps, thresholds etc. but this must be at the discretion of the operator who must assess all likely conditions.

• One operator can safely load/unload a patient using the Chair from a vehicle fitted with a ramp, automatic tail-lift or other approved mechanical lifting mechanism.

• Circumstance such as access, weather conditions or patient weight may require additional help when operating on stairs; this must be assessed at the time however the following is a suggested configuration:

1. Placement techniques, heavy patients, rough terrain or unusual circumstances may require more additional help; this must be assessed at the time however the following is a suggested configuration.

2. Keep the load limit of the Chair in mind when working with a heavy patient. If it suspected that the load limit has been exceeded then a full inspection of the Chair should be conducted, preferably by FERNO Field care.

3. Operators must take extra precautions when operating the Chair on uneven ground. If the ground is too uneven it is recommended to transfer the patient by pulling the chair backwards. With the back of the patient facing the operator.

WARNING

Regular cleaning helps reduce the risk of transmitting disease and enables the equipment to function at its optimum

4. Extra precautions should be taken to prevent the Chair from slipping in adverse, wet or icy weather conditions.

5. Stay with the patient at all times and always use the patient restraints provided.

6. Always follow good lifting procedures as specified in current manual handling regulations and your appropriate training manual.

7. Follow standard emergency patient-handling procedures when operating the Chair.

8. Medical advice is beyond the parameters of the manual. It is the user's responsibility to ensure safe practices for both themselves and the patient.

WARNING

Improper use can cause injury! Use the Chair only for the purpose described in this manual.

5.2. Moving the Chair

• Place the patient on the Chair using local protocols for patient handling. Ensure that the patient is seated with their back fully against the back panel and their head securely positioned within the head pad (see page 6 for fitting of Head Pad) before using the lap restraint.

• Chest and foot restraints must be applied before moving the Chair.

• When moving patients always inform them of what you are going to do.

5.3. Using the Chair on Level Ground



• Before rolling the Chair on level ground, the head end operator adjusts the upper handle to the desired position for comfortable use. (Rear locking lift handle is intended only for over stairs transport, not for chair turning on level ground).

• Inform the patient so that they are aware about what is happening.

• Two castor wheels are affixed to the foot end frame. These assist with movement on level ground and mean that the Chair can be used without tilting backwards.

• The operator at the foot end of the Chair assists as required with obstacles like door frames or uneven surfaces. The lower handle can be used to overcome larger obstacles.

5.4. Gliding the Chair Downstairs



• It is strongly recommended that the following procedures be practiced with a chair loaded with dummy weights such as bagged builders sand or gravel.

• Using the Chair on stairs requires a minimum of two operators. Use additional help as needed to control the Chair.

• It is recommended that the two operators face each other when transporting a patient down stairs. If available a third person should assist by acting as a guide for the foot-end operator.

• Ensure the route down the stairs is clear of obstructions that could affect the safety of the operators or the function of the tracks (see page 10).

1. **Head-end operator** – fully extend the upper handle and ensure it is locked (see page 9).

2. **Head-end operator** – lower the track and ensure it is locked (see page 11).

3. **Foot-end operator** – take up position on the stairs facing the chair gripping the lower foot-end handle.

4. **Both operators** – working together, guide the Chair over the edge or the top step, allowing the track belts to engage the step.

5. **Both operators** – move the Chair slowly downwards until the tracks are resting on the top two or three steps.

This establishes the glide angle. Ensure both operators are comfortable before proceeding.



⚠ WARNING
Before positioning the patient on the Chair, always ensure the Chair is fully unfolded and that the safety rings are up and over each hinge bracket.
Avoid rapid Chair repositioning as this may result in patient injury. Support the Chair while adjusting handles.
Never leave a patient unattended in the Chair and always use restraints.

6. **To maintain the glide angle** –

• **Foot-end operator** – applies slight downward pressure on the extended upper handle.

• **Head-end operator** – applies slight upward lift to the foot-end lift handles.

7. **Both operators** – when chair reaches the bottom of the steps and rear wheels are on the floor, the foot end operator releases the foot-end handle, the head-end operator tips the Chair forward until all four wheels are on the floor.

8. If this is the end of the stairs then the **Head-end operator** can fold the track and lower the upper handle to a comfortable height.

Head end operator leans the Chair backwards until the patient's weight is on the transport wheels.



At this point, the lower handle is more easily accessible to the foot end operator. The Chair can then be rolled towards the foot of the stairs. Position both track assemblies squarely against the facing edge of the step. The operators can then move the Chair up the stairs by pushing (foot end operator) and pulling (head end operator).

The head end operator is responsible for guiding the Chair up the stairs whilst the foot end operator assists as required.

If there is a need to manoeuvre the Chair on stairs the head end operator may need to adjust the upper handle to a higher or lower position.

TO PAUSE DURING DESCENT

To pause during the descent, tilt the Chair forward just enough to allow the rear wheels to rest on the step while both operators hold the Chair in place. To continue down the stairs, tilt back the Chair and regain the glide angle (see step 5 above)

Up Stairs: Once the patient is seated on the Chair and restrained, adjust the upper handle into a position that is comfortable for your height. (See Section Upper Handle). The upper handle position can be changed to provide a better hand grip.

Always inform the patient about what is happening.

6. Care of the FAST Chair

6.1. Disinfection

Disinfect all surfaces of the Chair including the head pad and patient restraints. Follow disinfectant manufacturers' directions for use. Spray or wipe disinfectant onto the patient restraints. Do not immerse in disinfectant.


6.2. Cleaning

Clean the Chair and its components with a clean cloth or soft brush and warm soapy water. Rinse with warm water and air dry. Do not clean with products containing bleach or phenol. A stiff bristled brush can be used if necessary.

7. Using additional help

Operating the chair requires a minimum of two trained operators. Ferno recommends positioning a trained operator at each end of the chair, that operators and helpers face each other when transporting a patient up or down stairs, and that a third trained person "spot" for the lead operator. Follow all applicable local protocols for carrying chairs.

The trained operators should maintain control of the chair and operate the controls, and the designated lead operator should direct all helpers. The charts below show suggested placement for operators and helpers.

Direction	Rolling on Flat Surface
Two Operators and Two helpers	





WARNING

Helpers can cause injury or be injured. Maintain control of the chair, operate the controls, and direct all helpers.



Load Limit

Inspect the chair if the load limit has been exceeded (See *Inspecting the Chair,* page 37).

Direction	DESCEND	ASCEND
Two Operators and One helper		
Two Operators and Three helpers		

Key: O = Operator H = Helper S = Spotter P = Patient

8. Maintenance

The FAST Chair requires regular maintenance. Set up and follow your own maintenance schedule but use the following as a minimum.

More frequent maintenance will be required in situations of heavy use. If any damage is discovered, take the Chair out of service until repair or replacement can be made.

When using any product for maintenance, follow the manufacturers' guidelines and read the manufacturers' material safety data sheet.

Action	As Needed	Each Month
Disinfecting	•	
Cleaning	•	
Inspection	•	•
Lubrication		•

⚠ WARNING	⚠ WARNING
Improper maintenance can cause injury. Maintain the Chair only for the purpose described in this manual.	**Improper use can cause injury! Use the Chair only for the purpose described in this manual.**

9. Warranty and Service

Standard terms and conditions apply to all sales. A copy is available upon request. These contain full details of warrant terms and do not limit the statutory rights of the customer.

For service, maintenance and any questions regarding this, or any other FERNO product, please contact:

Ferno Slovakia s.r.o.

Bošáca 893, 913 07

Slovakia

Tel:+421(0)327708010

or your local distributor

Registered Office:

FERNO Slovakia s.r.o.

Bošáca 893, 913 07 Slovakia

Tel: + 421 (0) 32 7708 010

e-mail: sales.sk@ferno.com

As our policy is one of continuous development FERNO Slovakia s.r.o. reserves the right to change the specifications without notice.

TRAINING RECORD		
Date	**Name**	**Training Method**

MAINTENANCE RECORD		
Date	**Maintenance Performed**	**by**